



Módulo ISON 3**Light Source Test Report****Color Parameters:****Chromaticity Coordinate:** $x=0.3053$ $y=0.3262/u'=0.1937$ $v'=0.4657$ **Tc=6947K Dominant WL:Ld=488.4nm Purity=10.0% Centroid WL:538.0nm****Ratio:R=12.6% G=83.3% B=4.0% Peak WL:Lp=450.0nm HWL:21.0nm****Render Index:Ra=73.0****R1 =70 R2 =77 R3 =80 R4 =74 R5 =71 R6 =68 R7 =83****R8 =61 R9 =-32 R10=43 R11=70 R12=39 R13=71 R14=89 R15=65****Photo Parameters:****Flux: 74.260 lm Fe: 0.23051 W Efficacy:96.84 lm/W****LEVEL:OUT WHITE:ANSI_6500K****Electrical Parameters:****Luminaire: U=12.00V I=0.06390A P=0.7668W PF=1.000****Instrument Status:****Scan Range:400.0nm-800.0nm Interval:5.0nm[0]****Ip=30490(G=6,D=53)****REF=5843(R=4) %=-0.261%****PMT: 16.2 centigrade [21.8]****Product Type:TEST DE CALIBRACIÓN****Number:058****Temperature:25.3 deg****Test Operator:2017-11-06 11:19:44****Software:V2.00.100**