



Arae 6**Light Source Test Report****Color Parameters:**

Chromaticity Coordinate: $x=0.3771$ $y=0.3733$ $u'=0.2243$ $v'=0.4995$

Tc=4068K Dominant WL:Ld=579.2nm Purity=25.2% Centroid WL:568.0nm

Ratio:R=19.7% G=76.9% B=3.4% Peak WL:Lp=450.0nm HWL:21.4nm

Render Index:Ra=83.2

R1 =82 R2 =90 R3 =95 R4 =82 R5 =82 R6 =85 R7 =86

R8 =65 R9 =10 R10=75 R11=81 R12=60 R13=84 R14=97 R15=76

Photo Parameters:

Flux: 989.58 lm Fe: 3.0079 W Efficacy:109.7 lm/W

LEVEL:OUT WHITE:ANSI_4000K

Electrical Parameters:

Luminaire: U=24.00V I=0.3758A P=9.019W PF=1.000

Instrument Status:

Scan Range:400.0nm-800.0nm Interval:5.0nm[0]

REF=7830(R=3)

%=-0.013%

Ip=54600(G=5,D=51)

PMT: 15.4 centigrade [21.3]

Product Type:TEST DE CALIBRACIÓN

Number:057

Temperature:25.3 deg

Test Operator:2017-11-06 11:19:44

Software:V2.00.100