



Alfa 9**Light Source Test Report****Color Parameters:**

Chromaticity Coordinate: $x=0.4562$ $y=0.4142$ $u'=0.2585$ $v'=0.5282$

Tc=2780K Dominant WL:Ld=583.3nm Purity=61.3% Centroid WL:593.0nm

Ratio:R=26.3% G=71.4% B=2.4% Peak WL:Lp=605.0nm HWL:112.4nm

Render Index:Ra=81.6

R1 =81 R2 =93 R3 =93 R4 =78 R5 =81 R6 =93 R7 =80

R8 =54 R9 =2 R10=85 R11=78 R12=75 R13=84 R14=96 R15=72

Photo Parameters:

Flux: 1199.6 lm Fe: 3.6089 W Efficacy:85.35 lm/W

Electrical Parameters:

Luminaire: U=24.00V I=0.5856A P=14.05W PF=1.000

Instrument Status:

Scan Range:400.0nm-800.0nm Interval:5.0nm[0]

REF=9471(R=3)

%=-0.043%

Ip=29068(G=5,D=51)

PMT: 16.9 centigrade [23.5]

Product Type:TEST DE CALIBRACIÓN

Number:057

Temperature:25.3 deg

Test Operator:2017-11-06 11:19:44

Software:V2.00.100