



Alfa +2**Light Source Test Report****Color Parameters:**

Chromaticity Coordinate: $x=0.3830$ $y=0.3810/u'=0.2251$ $v'=0.5038$

Tc=3962K Dominant WL:Ld=578.6nm Purity=29.3% Centroid WL:587.0nm

Ratio:R=21.5% G=74.4% B=4.2% Peak WL:Lp=455.0nm HWL:33.5nm

Render Index:Ra=96.0

R1 =97 R2 =99 R3 =98 R4 =93 R5 =94 R6 =96 R7 =95

R8 =94 R9 =90 R10=96 R11=94 R12=71 R13=98 R14=99 R15=95

Photo Parameters:

Flux: 1033.1 lm Fe: 3.7896 W Efficacy:74.09 lm/W

LEVEL:OUT WHITE:ANSI_4000K

Electrical Parameters:

Luminaire: U=24.00V I=0.5810A P=13.94W PF=1.000

Instrument Status:

Scan Range:400.0nm-800.0nm Interval:5.0nm[0]

REF=8170(R=3)

%=-0.025%

Ip=51707(G=5,D=52)

PMT: 17.7 centigrade [24.0]

Product Type:TEST DE CALIBRACIÓN

Number:057

Temperature:25.3 deg

Test Operator:2017-11-06 11:19:44

Software:V2.00.100