



Atik 2+**Light Source Test Report****Color Parameters:****Chromaticity Coordinate: x=0.4639 y=0.4065/u'=0.2670 v'=0.5264****Tc=2613K Dominant WL:Ld=585.3nm Purity=61.3% Centroid WL:621.0nm****Ratio:R=29.6% G=67.8% B=2.5% Peak WL:Lp=645.0nm HWL:146.7nm****Render Index:Ra=97.2****R1 =98 R2 =97 R3 =99 R4 =99 R5 =97 R6 =94 R7 =96****R8 =97 R9 =96 R10=96 R11=97 R12=85 R13=97 R14=99 R15=99****Photo Parameters:****Flux: 815.80 lm Fe: 3.3367 W Efficacy:63.18 lm/W****LEVEL:OUT WHITE:ANSI_2700K****Electrical Parameters:****Luminaire: U=24.00V I=0.5380A P=12.91W PF=1.000****Instrument Status:****Scan Range:400.0nm-800.0nm Interval:5.0nm[0]****Ip=19905(G=5,D=52)****REF=6472(R=3) %=-0.063%****PMT: 18.8 centigrade [25.0]****Product Type:TEST DE CALIBRACIÓN****Number:057****Temperature:25.3 deg****Test Operator:2017-11-06 11:19:44****Software:V2.00.100**